

**GOVERNMENT OF ASSAM
OFFICE OF THE DIRECTORATE OF MEDICAL EDUCATION, ASSAM SIX MILE,
KHANAPARA, GUWAHATI-22**

No. DME.E-674645-9

Dated ↓

Corrigendum-2

With reference to the IFB No: **AHIDMS/JICA/P&C/Equipment/RAD/2025/07 for Supply, Installation and Commissioning of Diagnostic & Imaging Equipment for AMCH, GMCH, LMCH, DMCH, JMCH, FAAMCH and TMCH in Assam UNDER JICA FUNDED PROJECT under Assam Health System Strengthening Project (AHSSP)** certain amendments have been incorporated in the IFB document. Bidders are requested to take note of these amendments prior to the submission of bid.

The following corrigendum in the subject Tender issued is hereby authorized: -

Lot-1 : SPECIFICATIONS OF DIGITAL RADIOGRAPHY 1000mA

<u>Existing</u>	<u>Corrigendum</u>
L.Additional Requirements	Bidders will have to arrange for functional demonstration of the offered equipment on the notified date. The evaluation of the spec indicator will be made based on the reports of the functional demonstration of the equipment. The functional demonstration (i.e. onsite /offsite) of the equipment's purely at the discretion of the Technical Bid Evaluation Committee and its input shall be treated as only corroborative in nature and will not be a substitute for technical evaluation of the document submitted along with the bid. The decision of the Technical Committee in this regard will be final. A bidder will be considered technically qualified if, Comply with above

Lot-2 : SPECIFICATIONS OF DIGITAL MAMMOGRAPHY

<u>Existing</u>	<u>Corrigendum</u>
F) STAND ASSEMBLY	
- Anti scatter grid: motorized movement, grid ratio 5:1, 35lp/cm with Carbon fiber interspaced material.	Anti-scatter grid: motorized movement, grid ratio 5:1, 31lp/cm or more with Carbon fiber interspaced material or equivalent technology

All other terms and conditions of the Bid Document shall remain unchanged.

Director of Medical Education, Assam