

**GOVERNMENT OF ASSAM
OFFICE OF THE DIRECTORATE OF MEDICAL EDUCATION, ASSAM
SIXMILE, KHANAPARA, GUWAHATI-22**

No.DME.E-674645-8

Dated ↓

Corrigendum-2

With reference to the IFB No: AHIDMS/JICA/P&C/Equipment/Imaging/2025/08 DATED 09-10-2025 FOR Supply, Installation and Commissioning of Medical Imaging Equipment for AMCH, GMCH, SMCH, LMCH, DMCH, JMCH, FAAMCH and TMCH in Assam UNDER JICA FUNDED PROJECT under Assam Health System Strengthening Project (AHSSP) certain amendments have been incorporated in the IFB document. Bidders are requested to take the note of these amendments prior to the submission of bid.

The following corrigendum in the subject Tender issued is hereby authorized: -

Lot-1: Technical Specification of Single-loader CR System with Dry imager	
Existing	Corrigendum
Image Reading (CR Reader /Digitizer)	
3) The CR reader / digitizer should be able to process at least 100 image plates/hr or more.	The CR reader / digitizer should be able to process at least 60 image plates/hr or more for largest size of cassette.

All other terms and conditions of the Bid Document shall remain unchanged.

Director of Medical Education, Assam