

GOVERNMENT OF ASSAM
OFFICE OF THE DIRECTOR OF MEDICAL EDUCATION, ASSAM SIXMILE,
KHANAPARA, GUWAHATI-22

No. DME.E- 744306/12

Dated ↓

Corrigendum-1

With reference to the IFB No: AHIDMS/JICA/P&C/Equipment/Neuro-Micro/2026/RE/04; DATED 20-07-2026 invited for SUPPLY, INSTALLATION AND COMMISSIONING OF NEURO SURGERY OPERATING MICROSCOPE AT SMCH, Silchar in ASSAM, under Assam Health System Strengthening Project (AHSSP) the following corrigendum is hereby issued based on the decisions taken during the technical review meeting. This corrigendum shall be part of the tender document.

The following corrigendum in the subject Tender issued is hereby authorized: -

Sl. No	Technical Specification	Recommendation of Committee
B	BINOCULAR TUBE:	
1	180 degree range tiltable binocular tube with focal length = 160 mm or higher . Should Graduate knob for continuous adjustment of interpupillary distance from 55 mm to 75 mm	360 degree Tilttable, rotatable, length adjustable binocular tube with focal length 160 mm or higher. Should have incremental stop loss in the tube in its desired positions.
D	Eye piece Lens	
1	Stereo Co Observation attachment with two joints with side changer.	Points to be deleted
G	FLOOR STAND:	
1	Counterweight balance technology	Counterweight balance technology or suspension system.
2	Rollable floor stand on base with lockable castors, carrier and swivel arms with Large reach of 1.7 m or higher, weight carrying capacity of at least 12 kg or more.	Rollable floor stand with lockable castors, carriers and swivel arms with Large reach of 1.7 m or higher, weight carrying capacity of at least 12 kg or more.
H	INTEGRATED DIGITAL VIDEO CAMERA SYSTEM:	
1	Advanced digital 3 chip Full HD Video camera (1920 x 1080p) should be attached to supply output to the stand mounted colour LCD screen and for recording purpose.	Advanced digital 3 chip Full HD Video Camera- Which should not be visible from outside (should be attached to supply output to the stand mounted colour LCD screen and for recording purpose).

L	ADDITIONAL FEATURES: (Price should be quoted separately)	
1	The microscope should be capable to superimpose the image of CT scan, MRI and Endoscopy (image guided neuro surgery). Microscope should work with Neuro navigation system available with the hospital.	The microscope should be capable to superimpose the image of CT scan, MRI and Endoscopy (image guided neuro surgery)- Optional The microscope should be capable to work with Neuro Navigation system available with the Hospital.
	Additional Point-1:- Service support in Assam	The bidder/OEM shall have an authorized service support facility in the state of Assam. The bidder shall submit the escalation matrix, including the contact details of the Service Engineer(s) based in Assam. In the event of a service complaint, the service engineer shall physically visit the consignee site within 48 hours from the time of complaint registration.
	Additional Point-2:- Preventive Maintenance (PM)	The bidder/ OEM shall carry out preventive maintenance (PM) visit atleast 3 (Three) times in a tear during the warranty and comprehensive maintenance contract (CMC) period. A Preventive Maintenance scheduled shall be shared with the consignee and each visit shall be documented and acknowledged by the user/consignees.

All other terms and conditions of the Bid Document shall remain unchanged.

Director of Medical Education, Assam